

Rapid Precision Testing Laboratories

P.O. Box 1342
Cordova TN 38088-1342
Telephone: (901) 386-0175
Fax: (901) 386-7218
e-mail: RPTL@aol.com

Robert M. Sayre, Ph.D.

August 29, 2001

To: Mr. Garry Hollar

From: Robert M. Sayre, Ph.D.

Subject: Review Chairside Whitening System

Dear Mr. Hollar:

I have reviewed the Chairside Whitening System Protocol. I believe that the risks to the patients have been minimized for this new therapy system for whitening teeth. I also believe that while risks for the dentist and dental technician is minimal, it too has been further minimized by the use of protective eyewear.


I would suggest that two additional groups of potentially light sensitive individuals also be definitely excluded: patients on PUVA (Psoralen + UV Radiation) or other photochemotherapy for any condition. Such patients might have special sensitivity to the wavelengths being used in teeth whitening.

The other population I would definitely exclude would be anyone with melanoma or other light sensitive condition. This most virulent form of skin cancer is very poorly understood at best and many consider UV a risk factor.

I would also suggest that patients using any of the list of light sensitive drugs provided by the FDA to tanning salons should be excluded from the initial study.

Personally I believe most of the drugs on the FDA's list are not a problem, but I recommend your company will be well served by formally excluding them from the initial studies with this new dental whitening system.

Thank you.



Robert M. Sayre, Ph.D.

Attached : RMS CV